## BNPM/LTE/ROTAMETE R/0231/2020-21 Limited Tender Enquiry for Supply of Rotameter at BNPM, Mysuru ADDENDUM NO. 1

## **ADDENDUM No. 1, DATED 10.07.2020**

## 1.0 SCOPE OF THIS ADDENDUM

- 1.1 This corrigendum dated 10.07.2020, is issued for,
  - (a) Clarifications to queries sought by bidders.
    - a. Fluid: Wet Strength Resin (WSR).
    - b. Viscosity: 10 to 30 centipoise at 30 degree Celsius.
    - c. Density: **1.10-1.15 g/cc.**
    - d. Operating pressure: **0-5 bar.**
    - e. Operating temperature: 20-60 degree Celsius.
    - f. Flow Range: **0-15 Liters per minute (lpm).**
    - g. Size of end connections (hose size): ½ inch threaded type.