

BNPM/LTE/ROTARY EVAPORATOR/0458/ 2020-21	<b>BANK NOTE PAPER MILL INDIA PRIVATE LIMITED</b>	SHEET 1 OF 1
	<b>Tender for Supply, Installation &amp; Demonstration of Lab Rotary Evaporator at BNPM Plant, Mysuru</b> <b>CORRIGENDUM NO. 2</b>	

**CORRIGENDUM No. 2, DATED 15.09.2020**

**FOR**

**TENDER NO. BNPM/LTE/ROTARY EVAPORATOR/0458/2020-21 Dated**  
**18.08.2020**

**TENDER FOR SUPPLY, INSTALLATION &  
DEMONSTRATION OF LAB ROTARY EVAPORATOR  
AT BNPM PLANT, MYSURU**



BNPM/LTE/ROTARY EVAPORATOR/0458/ 2020-21	<b>BANK NOTE PAPER MILL INDIA PRIVATE LIMITED</b>	SHEET 1 OF 1
	<b>Tender for Supply, Installation &amp; Demonstration of Lab Rotary Evaporator at BNPM Plant, Mysuru</b> <b>CORRIGENDUM NO. 2</b>	

**1.0 SCOPE OF THIS CORRIGENDUM**

1.1 This corrigendum dated 15.09.2020, is issued for,

(a) The tender due date extension for submission of tender floated on 18.08.2020.

(b) Clarification to S No. 5 of Section VII (Technical Specifications) of tender.

1.2 Except for details mentioned herein, all other details contained in the tender no. BNPM/LTE/ROTARY EVAPORATOR/0458/2020-21 dated 18.08.2020, shall remain applicable and unchanged.

**2.0 REVISION TO THE TENDER:-**

2.1 Clause no. 1.1 (a) is revised as,

Closing date and time for receipt of tenders	25-09-2020 - 1500 Hours
Date and time of opening of Techno-commercial bid	25-09-2020 - 1600 Hours

2.2 Clause no. 1.1 (b) is revised as,

Section VII (Technical Specifications) of tender:-

S No	Parameters	Required Specification (as per tender)	Revised Specification (May be revised as)
5.	Temperature Range	Ambient to 180°C (Tolerance: ±5°C).	Ambient to 180°C • Lower limit: Ambient ±5°C. • Upper limit: 180°C or above.

