

BNPM/LTE/STROBOS COPE/0965/2020-21	BANK NOTE PAPER MILL INDIA PRIVATE LIMITED	SHEET 1 OF 1
	Tender for Supply of Handheld Stroboscope at BNPM Plant, Mysuru. CORRIGENDUM NO. 3	

**CORRIGENDUM No. 3, DATED 19.02.2021**

**FOR**

**TENDER NO. BNPM/LTE/STROBOSCOPE/0965/2020-21 Dated**  
**08.01.2021**

**TENDER FOR SUPPLY OF HANDHELD  
STROBOSCOPE AT BNPM PLANT, MYSURU**



BNPM/LTE/STROBOSCOPE/0965/2020-21	<b>BANK NOTE PAPER MILL INDIA PRIVATE LIMITED</b>	SHEET 1 OF 1
	<b>Tender for Supply of Handheld Stroboscope at BNPM Plant, Mysuru.</b> <b>CORRIGENDUM NO. 3</b>	

## 1.0 **SCOPE OF THIS CORRIGENDUM**

1.1 This corrigendum dated 19.02.2021, is issued for,

- (a) The tender due date extension for submission of tender floated on 08.01.2021.
- (b) Inserting clauses (VI) & (VII) in Section IX – Qualification/Eligibility Criteria.
- (c) Clarifications in Section VII - Technical Specifications.

1.2 Except for details mentioned herein, all other details contained in the tender no. BNPM/LTE/STROBOSCOPE/0965/2020-21 Dated 08.01.2021, shall remain applicable and unchanged.

## 2.0 **REVISION TO THE TENDER:-**

2.1 Clause no. 1.1 (a) is revised as

Closing date and time for receipt of tenders	02-03-2021 - 1500 Hours
Date and time of opening of Techno-commercial bid	02-03-2021 - 1600 Hours

2.2 Clause no. 1.1 (b) as,

VI) A bidder is permitted to procure raw material, components, sub-assemblies etc. from the vendors from countries which shares a land border with India. Such vendors will not be required to be registered with the competent authority, as it is not regarded as sub-contracting.

VII) However, in case a bidder has proposed to supply finished goods procured directly/indirectly from the vendors from the countries sharing land border with India, such vendor will be required to be registered with the competent authority.

2.3 Clause no. 1.1 (c) is as below,

S No	Tender Clause	Bidders Query	BNPM Remarks
1.	<b>Section VII</b> #10.4: UV Illumination – Yes.	Only UV can be provided without illumination	UV without illumination is acceptable.



BNPM/LTE/STROBOS COPE/0965/2020-21	<b>BANK NOTE PAPER MILL INDIA PRIVATE LIMITED</b>	SHEET 1 OF 1
	<b>Tender for Supply of Handheld Stroboscope at BNPM Plant, Mysuru.</b> <b>CORRIGENDUM NO. 3</b>	

2.	<b>Section VII</b> #10.5: Splash Proof – Yes #10.6: Ex-proof – Yes #10.7: Type of Protection – IP50 (or) above.	IP54 rating shall be provided	IP54 rating is acceptable.
3.	<b>Additional information sought by bidder</b>	Wavelength is not specified in technical specifications. 365nm & 385nm are available.	Both 365nm & 385nm wavelengths are acceptable.

