

<u>BNPM/NCB/486/2024</u> <u>-25 dated 20.01.2025</u>	BANK NOTE PAPER MILL INDIA PRIVATE LIMITED	SHEET 1 of 2
	SUPPLY, INSTALLATION AND 5 YEARS OF CAMC/AMC OF ANAESTHESIA WORK STATION WITH VENTILATOR, MONITOR AND 2 VAPORIZERS AT CHIGATERI DISTRICT HOSPITAL, DAVANAGERE, KARNATAKA UNDER CSR ACTIVITY OF BNPM PLANT, MYSORE	

CORRIGENDUM No. 1, DATED 11.02.2025

FOR

E-TENDER NO. BNPM/NCB/486/2024-25 dated 20.01.2025

1.0 SCOPE OF THIS CORRIGENDUM

A. Amendment in technical specification:

S. No.	Clause	Did read	Read as
1.	Section-VII – Technical Specifications, page 43 of 76, point no. B, 1, (iii).	Display:- It should have a high contrast color TFT Display	Display:- It should have a high contrast color TFT touch display size min. 12”.
2.	Section-VII – Technical Specifications, page 44 of 76, point no. B, 6.	Fresh Gas De-coupling:- The unit should have Fresh Gas De-coupling or Equivalent mechanism.	Fresh Gas Decoupling/ Compensation: The unit should have Fresh Gas De-coupling or Fresh gas Compensation or Equivalent Mechanism
3.	Section-VII – Technical Specifications, page 44 of 76, point no. C, 1, (iv).	other data for 480 hrs	Other data / trends between 120 hrs. – 168 hrs
4.	Section-VII – Technical Specifications, page 44 of 76, point no. C, 1, (v).	2000 groups of alarm events	1000 groups of alarm events
5.	Section-VIII – Quality Control requirements, page 47 of 76, Sl no. 13, B, (iii).	Display:- It should have a high contrast color TFT Display	Display:- It should have a high contrast color TFT touch display size min. 12”.
6.	Section-VIII – Quality Control requirements, page 48 of 76, Sl no. 18.	Fresh Gas De-coupling:- The unit should have Fresh Gas De-coupling or Equivalent mechanism.	Fresh Gas Decoupling/ Compensation: The unit should have Fresh Gas De-coupling or Fresh gas Compensation or Equivalent Mechanism.

<u>BNPM/NCB/486/2024</u> <u>-25 dated 20.01.2025</u>	BANK NOTE PAPER MILL INDIA PRIVATE LIMITED	SHEET 2 of 2
	SUPPLY, INSTALLATION AND 5 YEARS OF CAMC/AMC OF ANAESTHESIA WORK STATION WITH VENTILATOR, MONITOR AND 2 VAPORIZERS AT CHIGATERI DISTRICT HOSPITAL, DAVANAGERE, KARNATAKA UNDER CSR ACTIVITY OF BNPM PLANT, MYSORE	

7.	Section-VIII – Quality Control requirements, page 48 of 76, Sl no. 20, iv.	Other data for 480 hrs	Other data / trends between 120 hrs. – 168 hrs
8.	Section-VIII – Quality Control requirements, page 48 of 76, Sl no. 20, (v).	2000 groups of alarm events	1000 groups of alarm events

B. Bid Submission due date is extended.

<u>S. No.</u>	<u>Previous Provision</u>	<u>Action/ Clarification incorporated in this corrigendum</u>
01	Due date & time for bid submission: 11.02.2025 & 1100 Hrs	Due date & time for bid submission: 18.02.2025 & 1100 Hrs
02	Opening date and time for tenders: 11.02.2025 & 1130 Hrs	Opening date and time for tenders: 18.02.2025 & 1130 Hrs

Except for details mentioned herein, all other details contained in the tender no. BNPM/NCB/486/2024-25 dated 20.01.2025, shall remain applicable and unchanged.