

<u>BNPM/NCB/VENTILAT</u> <u>OR/ 0583/2021-22</u> <u>Dated: 31.12.2021</u>	BANK NOTE PAPER MILL INDIA PRIVATE LIMITED	SHEET 1 OF 1
	Tender for supply and installation of ICU ventilators at K.R.HOSPITAL, Mysore under CSR activity of BNPM plant, Mysore CORRIGENDUM NO. 1	

CORRIGENDUM No. 1, DATED 28.01.2022

FOR

TENDER NO. BNPM/NCB/VENTILATOR/ 0583/2021-22

Dated: 31.12.2021

Tender for supply and installation of ICU ventilators at K.R.HOSPITAL, Mysore under
CSR activity of BNPM plant, Mysore



<u>BNPM/NCB/VENTILATOR/ 0583/2021-22</u> <u>Dated: 31.12.2021</u>	BANK NOTE PAPER MILL INDIA PRIVATE LIMITED	SHEET 1 OF 1
	Tender for supply and installation of ICU ventilators at K.R.HOSPITAL, Mysore under CSR activity of BNPM plant, Mysore CORRIGENDUM NO. 1	

1.0 **SCOPE OF THIS CORRIGENDUM**

1.1 This corrigendum dated 28.01.2022, is issued for,

- (a) The tender due date extension for submission of tender floated on 31.12.2021.
- (b) Clarification to the tender

1.2 Except for details mentioned herein, all other details contained in the tender no. BNPM/NCB/VENTILATOR/ 0583/2021-22 dated: 31.12.2021, shall remain applicable and unchanged.

1.3 **REVISION TO THE TENDER**

Clause no. 1.0, NIT, Sec – I is revised as

Closing date and time for receipt of tenders	04-02-2022 - 1100 Hours
Date and time of opening of Techno-commercial bid	04-02-2022 - 1200 Hours

Clarification to the tender:

Sr. No.	Attribute	Description
1.	Reference Clause	Sec – VII, Technical specification , Sr No 3
	Tender Description	Ventilator capable to deliver tidal volume from 100 ml to 2000 ml in Volume control mode and min. 50 ml in Pressure Control mode.
	Amendment requested	Ventilator capable to deliver tidal volume from 20 ml to 2000 ml in Volume control mode and min. Remarks: Many company can offer this tidal volume range to cover



<u>BNPM/NCB/VENTILAT</u> <u>OR/ 0583/2021-22</u> <u>Dated: 31.12.2021</u>	BANK NOTE PAPER MILL INDIA PRIVATE LIMITED	SHEET 1 OF 1
	Tender for supply and installation of ICU ventilators at K.R.HOSPITAL, Mysore under CSR activity of BNPM plant, Mysore CORRIGENDUM NO. 1	

Sr. No.	Attribute	Description
		pediatric and adult patients. 20 - 2000 ml also offers adequate safety margin to use it even in small pediatric patients.
	Clarified as	No change in technical specification.
2.	Reference Clause	Sec – VII, Technical specification , Sr No 6
	Tender Description	Non-Invasive ventilation modes – Pressure Control & Pressure Support with minimum Leak compensation upto 60 Ltrs./Min. and more
	Bidders Query	Dedicated NIV mode for mask ventilation with leak compensation. Remarks: The dedicated NIV mode adapts to patient's breath pattern and supports spontaneous efforts while takes full control in Apnea situation.
	Clarified as	No change in technical specification.
3.	Reference Clause	Sec – VII, Technical specification , Sr No 8
	Tender Description	Should have facility to monitor exhaustive lung mechanics including Plateau Pressure, End Expiratory Flow, Work of Breathing Ventilator & Patient.
	Bidders Query	Should have facility to monitor exhaustive lung mechanics including Plateau Pressure, End Expiratory Flow, time constant, auto patient detect.



<u>BNPM/NCB/VENTILAT</u> <u>OR/ 0583/2021-22</u> <u>Dated: 31.12.2021</u>	BANK NOTE PAPER MILL INDIA PRIVATE LIMITED	SHEET 1 OF 1
	Tender for supply and installation of ICU ventilators at K.R.HOSPITAL, Mysore under CSR activity of BNPM plant, Mysore CORRIGENDUM NO. 1	

Sr. No.	Attribute	Description
		Remarks: Time constant is an actionable data to select expiratory time and auto patient detect alarm helps to avoid unforeseen errors leading to patient suffocation due to ventilator not being turned ON from Standby.
	Clarified as	No change in technical specification.
4.	Reference Clause	Sec – VII, Technical specification , Sr No 13
	Tender Description	Ventilator display size min. 12” with direct access to vital parameters controls such as FiO2, PEEP, Rate, Vt, / PC directly.
	Bidders Query	Ventilator display size min. 15” with direct access to vital parameters controls such as FiO2, PEEP, Rate, Vt, / PC directly. Remarks: Most modern ventilators have 15" screen to display more clinical data and large font for better visibility from a distance.
	Clarified as	No change in technical specification.
5.	Reference Clause	Sec – VII, Technical specification , Sr No 16
	Tender Description	Should have the facility for display of 24 hour trend with zoom in, also recall for alarm & event.
	Bidders Query	Should have the facility for display of 72 hour trend with zoom in, also recall for alarm & event. Remarks: Literature suggest any overt event should be retrospectively



<u>BNPM/NCB/VENTILAT</u> <u>OR/ 0583/2021-22</u> <u>Dated: 31.12.2021</u>	BANK NOTE PAPER MILL INDIA PRIVATE LIMITED Tender for supply and installation of ICU ventilators at K.R.HOSPITAL, Mysore under CSR activity of BNPM plant, Mysore CORRIGENDUM NO. 1	SHEET 1 OF 1
---	--	--------------

Sr. No.	Attribute	Description
		troubleshooted for past 3 days for early warning signs. This will help to establish effective and optimal ICU protocol to minimize patient related avoidable events.
	Clarified as	No change in technical specification.
6.	Reference Clause	Nil
	Tender Description	Nil
	Bidders Query	Integrated ETCO2 module - mainstream / side stream. Remarks: Early identification of patient disconnect and other critical clinical events.
	Clarified as	No change in technical specification.

