

**BANK NOTE PAPER MILL INDIA PRIVATE LIMITED,**  
**CIN: U21090KA2010PTC055475**  
**CORPORATE OFFICE, MYSORE**  
[scm@bnpmindia.com](mailto:scm@bnpmindia.com); [ssharma@bnpmindia.com](mailto:ssharma@bnpmindia.com)  
**Ph: 0821 2401 179 / 180**

**ENQUIRY NO. BNPM/ENQ/Lab Instruments/47/2018-19**  
**Enquiry issuing Date: 23.04.2018**  
**Closing date: 25.04.2018**

**1. Scope of work:**

Supply of Lab Instruments

**2. Terms & Conditions:**

S. No.	Particulars	Quantity
1.	Payment Terms	100% on receipt and acceptance of goods by the consignee at destination and on production of all required documents by the supplier. <b><u>Warranty certificate (if applicable) and NEFT/RTGS details shall be furnished along with invoice.</u></b>
2.	Price	Price should be inclusive of taxes,
3.	Freight	Shall be included in quoted price
4.	Packing & Forwarding	Shall be included in quoted price
5.	Delivery Terms	DAP, Bank Note Paper Mill India Private Limited, Mysore
6.	Warranty Period	.....
7.	Warranty Certificate	Applicable / Not Applicable
8.	Delivery Period	.....
9.	Material to be delivered at	Engineering Stores Bank Note Paper Mill India Private Limited Note Mudran Nagar Mysore 570 003
10.	Contact person in Engineering Stores	Mr. Ashutosh Kumar 0821 2401 486 Mr. Sunil Kumar 0821 2401 485



**3. Price Bid Format :**

S. No.	Description	Qty.	UOM	Unit rate	Freight, P&F Charges	GST@..	Unit rate including GST & Freight, P&F Charges	Grand total inclusive of GST and Freight, P&F Charges (DAP, BNPM, Mysore)
1.	Dissolved Oxygen meter	01	Nos					
2.	COD Digester	01	Nos					
3.	Turbidity meter	01	Nos					
4.	TDS Meter Electrode	01	Nos					
<b>Total price inclusive of all tax, freight , P&amp;F charges</b>								
<b>Technical Specifications for above mentioned items are attached as Annexure 1</b>								

We hereby confirm that

1. We accept all terms & conditions mentioned in the enquiry.
2. Price quoted is inclusive of all taxes, P&F, freight etc. on DAP, BNPM, Mysore basis.

Signature of bidder: .....

Name of the Firm: .....

Seal of the firm: .....

GST No: .....



Annexure - 01

Technical specification for Dissolved Oxygen (DO) meter			Remarks
1	Type	Benchtop / laboratory	
2	DO range	0.00 to 20.00 PPM(mg/L)	
3	Resolution	0.01ppm(mg/L)	
4	Accuracy	±1% of reading	
5	Calibration	one or two at 0%(in standard solution) and 100%(in air)	
6	Temperature compensation	Automatic	
7	Salinity compensation	Automatic; 0 to 40g/L(with 1g/L resolution)	
8	Altitude and barometric pressure compensation	Automatic	
9	Temperature range	(-)20.0 to (+)120.0°C	
10	Temperature Resolution	0.1°C	
11	Temperature Accuracy	±0.2°C	
12	Ambeint Temperature	0 to 50°C	
13	Power supply	230V AC	
14	Warranty	2 years from the date of installation	
15	Suitable for	Water treatment and Effluent treatment plant. Effluent - Pulping Plant & Paper Machines	
16	Instrument housing, keys and display	Water resistant rubberised	
17	Make	Hanna Instruments/ Cole-Parmer/ Suntronics Inc/Hach	

Technical specification for COD meter			Remarks
1	Type	Benchtop / Laboratory use	
2	Samples holding capacity	Min 15 number of vials	
3	COD tubes	Min 15 number of vials	
4	COD Condensors	Min 15 number of vials	
5	COD Stand	Should be capable of holding 15 COD vials	
6	Temperature Range	(-) 10 to (+) 160 °C	
7	Temperature Accuracy	±2°C (@ 25°C)	
8	Temperature Stability	±0.5°C	
9	Power supply	230V AC, 50Hz	
10	Ambeint Temperature	5 to 50°C	
11	Warranty	2 years from the date of installation	
12	Suitable for	Water treatment and Effluent treatment plant. Effluent - Pulping Plant & Paper Machines	
13	Instrument housing, keys and display	Water resistant rubberised	
14	Make	Hanna Instruments/ Cole-Parmer/Hach/ Suntronics Inc	

Technical specification for Turbidity meter			Remarks
1	Type	Benchtop / Laboratory use	
2	Range	0-4000 NTU	
3	Principle of operation	Nephelometric & shall be compliant to ISO 7027	
4	Repeatability	$\pm 1\%$ of reading or $\pm 0.01$ NTU	
5	Accuracy	$\pm 2\%$ of reading plus 0.01 NTU from 0 - 1000 NTU $\pm 5\%$ of reading from 1000 - 4000 NTU	
6	Response time	Less than 7 seconds	
7	Sample size	30ml	
8	Operating temperature	0-50°C	
9	Power supply	230V AC, 50Hz	
10	Warranty	2 years from the date of installation	
11	Suitable for	Water treatment and Effluent treatment plant. Effluent Pulping Plant & Paper Machines	
12	Instrument housing, keys and display	Water resistant rubberised	
13	Make	Hanna Instruments/ Cole-Parmer/Hach/ Suntronic Inc	