Ph no. 0821-2401144; Email: sanjeev.kunal@bnpmindia.com

#### LIMITED TENDER ENQUIRY No. BNPM/LTE/CMHC/0187/2020-21

Issue Date: 06.06.2020 Extension No.3: 02.07.2020 Closing Date: 09.07.2020

### A. Scope of work:

Supply of below mentioned item:

Sl.no.	Description	Qty
1.	3Phase True RMS AC Power Clamp meter with Harmonics & Inrush Current measurement. Technical specifications at page no. 7	01 No.

#### **B.** Terms & Conditions:

- 1. Documents to be submitted:
  - a) Signed copy of this document (All the pages).
  - **b)** Filled price details as per the format provided below at point 17.
  - **c)** Technical details /datasheet for the enquired product (if applicable).
  - **d)** Micro / Small Enterprise (Certificate issued by MSME)/UAM/NSIC certificate (If applicable).
- 2. **Price:** Price shall be inclusive of P&F, Insurance, Freight and GST. Price shall remain fixed & no price variation shall be accepted till pendency of contract.
- 3. **Terms of Payment:** 100% within 30 days upon receipt and acceptance of goods at BNPM Plant, Mysuru and on production of all required documents by the supplier. Warranty certificate (if applicable) and NEFT/RTGS details shall be furnished along with the Original Invoice.
- Delivery Terms & Delivery Address: Delivery shall be made on F.O.R Basis, Bank Note Paper Mill India Pvt. Ltd., Mysore.
   Delivery Address: Engineering Stores, Bank Note Paper Mill India Pvt. Ltd., Note

Mudran Nagar, Mysore - 570003. Stores contact details: 0821-2401476/486.

5. **Taxes:** All Taxes shall be as applicable in GST regime.

Payment of CGST, SGST, IGST & UTGST: The suppliers are required to adhere the following procedure in order to honour the payment against CGST, SGST, IGST & UTGST in the invoice.

- i) An invoice issued by the vendor for goods or services or both as applicable should be in accordance with the provisions of Sec 31 of the CGST Act & should contain all the prescribed information's in accordance with Chapter VI of CGST rules 2017.
- ii) A debit note issued if any, by the vendor should be in accordance with the provisions of Sec 34 of the CGST Act.
- iii) The vendor should mandatorily upload the aforementioned documents in respective GSTR, details of outward supplies of goods or services as applicable within the prescribed time under GST Act.
- iv) The vendor should provide the relevant documents to confirm the tax charged on the invoice has been paid to the credit of government after adjusting the ITC if any.

Notwithstanding the above, the supplier should provide indemnification as follows:

In the event of non-compliances with respect to GST Act & Rules by the successful bidder, the purchaser is allowed to adjust the GST amount from retention amount (either in BG or in cash) held by the company. If no amount is available for recovery, the successful bidder will refund the GST liability within 10 days from the date of GST reversal in GSTRN.

The above requirements are mandatory to claim any GST liability, falling which, the GST liability will not be paid /reimbursed/accepted.

- 6. **Warranty:** The product supplied shall be under warranty for minimum 24 months from the date of receipt of material at BNPM Plant, Mysore. Warranty certificate to be provided on company letter head.
- 7. **Liquidated Damages (LD):** Liquidated damages shall be applicable at the rate of 0.5% per week or part thereof until actual delivery or performance, subject to a maximum deduction of 10% of the delayed good's /service's contract price.
- 8. **Delay in supplier's performance:** Time for and the date specified in the contract or as extended for the delivery of the stores shall be deemed to be the essence of the contract and the supplier shall deliver the goods and perform the services under the contract within the time schedule specified by BNPM in the contract.

Any delay attributable to the supplier in maintaining its contractual obligations towards delivery of goods and performance of services shall render the supplier liable to any or all the following sanctions besides any administrative action such as (a) Imposition of liquidity damages (as mentioned in point 7 of this enquiry document); (b) Termination of contract for default (as mentioned in point 10 of this enquiry document).

9. **Extension of Time:** If the contract is delayed in the progress of work by changes ordered in the work, or any clause which BNPM shall decide to justify the delay, then the time of completion shall be extended by a reasonable time.

If at any time during the currency of contract, the supplier encounters conditions hindering timely delivery of the goods and performance of services, the supplier shall promptly inform BNPM in writing about the same and its likely duration and make a request to BNPM for extension of the delivery schedule accordingly. On receiving the supplier's communication, BNPM shall examine the situation as soon as possible and, at its discretion, may agree to extend the delivery schedule, with or without liquidated damages for completion of supplier's contractual obligations by issuing an amendment to the contract.

- 10. **Rejection Replacement:** In case of supplies, materials not meeting our specification will be rejected outright and the rejected material shall be taken back within 7 days at the cost and risk of the supplier and replacement should be made within 15 days from the date of intimation. No payment shall be made for rejected item. If the material is not taken back within the stipulate period, BNPM reserves the right to dispose-off the material at the risk and expense of the vendor as per provision under Section IV: General conditions of contract.
- 11. **Risk Purchase Clause:** If the supplier fails to abide by the terms and conditions of the contract/agreement, or fails to supply the material as per the delivery schedule or any time repudiates the contract,

Procure the tendered item / render service from other agencies at the risk and cost of the supplier. The cost difference between the alternative arrangement and existing contract value wherein default has been made will be recovered from the supplier along with the other incidental charges.

In case of procurement through alternative sources, if procurement price is lower than the existing contract value wherein default has been made, in such case no benefit on this account will be passed on to the supplier.

- 12. **Repeat Orders:** Bidders may note that, quantity of maximum upto 50% of the total ordered quantity against this enquiry may be considered for repeat order within a period of 06 months from the date of completion of last supply made in the PO ordered against this enquiry with terms & conditions remaining same as per the original PO.
- 13. **Termination for Convenience (Foreclosure) Clause:** BNPM reserves the right to terminate the contract, in whole or in part for its (BNPM's) convenience, by serving written notice on the supplier at any time during the currency of the contract. The notice shall specify that the termination is for the convenience of BNPM. The notice shall also indicate inter-alia, the extent to which the supplier's

performance under the contract is terminated, and the date with effect from which such termination will become effective.

- 14. **Settlement of Disputes through Arbitration:** All disputes and differences of any kind whatsoever arising out of or in connection with the contract, whether during the progress of the work or after their completion except accepted matters shall be settled through arbitration process as per the Arbitration & Conciliation Act, 1996. The venue of arbitration shall be Mysore, Karnataka.
- 15. **Legal Jurisdiction:** The Courts of Mysore (Karnataka State) shall alone have jurisdiction to decide on any legal matter of dispute whatsoever arising out or in respect of the contract.
- 16. **Force Majeure:** In the event of any unforeseen event directly interfering with the supply of stores arising during the currency of the contract, such as war, hostilities, acts of the public enemy, civil commotion, sabotage, fires, floods, explosions, epidemics, quarantine restrictions, strikes, lockouts, or acts of God, the Contractor shall, within a week from the commencement thereof, notify the same in writing to the Purchaser with reasonable evidence thereof. Unless otherwise directed by BNPM in writing, the supplier shall continue to perform its obligations under the contract as far as reasonably practical, and shall seek all reasonable alternative means for performance not prevented by the

Force Majeure event. If the force majeure condition(s) mentioned above be in force for a period of 90 days or more at any time, either party shall have the option to terminate the contract on expiry of 90 days of commencement of such force majeure by giving 14 days' notice to the other party in writing. In case of such termination, no damages shall be claimed by either party against the other, save and except those which had occurred under any other clause of this contract prior to such termination.

17. **Rights of Rejection:** BNPMIPL reserves the right to reject any or all the applied bids without assigning any reason whatsoever. The enquiry can be rejected on national security grounds.

## 18. Price Bid Format:

Sl. No.	Material Description	UOM	Qty	Unit Price (INR)	P&F, Insurance & Freight Charges (INR)	GST@ ——— Amount (INR)	Unit price including P&F, Insurance, Freight & GST (INR)	Total price inclusive of P&F, Insurance, Freight & GST (F.O.R, BNPM, Mysuru) (INR)	Input GST Credit (INR) {GST amount * Total quantity}	Effective price inclusive of P&F, Insurance, Freight (F.O.R, BNPM, Mysuru) (INR)
A	В	С	D	Е	F	G	H=E+F+G	I=H*D	J=G*D	K=I-J
1.	3Phase True RMS AC Power Clamp meter with Harmonics & Inrush Current measurement. Technical specifications at page no. 7	No.	1							

Note: Evaluation shall be done on the effective price to decide L-1 bidder.

## We hereby confirm that

1.	We accept all the terms & conditions mentioned in the enquiry.
2.	Price quoted is inclusive of P&F, Insurance, Freight and GST on F.O.R Basis, BNPM
	Plant, Mysore.
3.	Bid validity: 30 days from the date of enquiry closing date.
4.	HSN Code:
5.	Delivery period: (Days / Weeks)
6.	Warranty Period: Months (Minimum 24 months required)
7.	Bank Details: Acc. No; Bank Name:;
	Branch name:; Branch Code:;
	IFSC:;
8.	MSME / NSIC status: (If yes, then supporting document to be
	submitted along with the offer)
	(Please fill above: MSI – For Micro Enterprises; SSI – For Small Enterprises; MED.SI –
	For Medium Enterprises; NSIC – For National Small Industries Corporation regd. firm)
Signat	rure of bidder:
Conta	ct Person:
Conta	ct Number:
Email	Id:
Name	of the Firm:
Seal of	f the firm:
GST N	0.:

# <u>Technical Specification for 3Phase True RMS AC Power Clamp meter with Harmonics & Inrush Current measurement:</u>

SL No.	3Ph	ase True RMS AC Power Clampmeter with Harmonics and inrush Current Measurement
1	Measurement Ranges	Current: 60 mA to 600 A
		Active power: 5W to 360 kW
2	Measurement line	Single-phase, Three-phase power system  Voltage, Current, Voltage/ current peak, Active/ reactive/ apparent
3	Measurement items	power, Power factor, Phase angle , Frequency, Simple
3	Weasurement items	Active Energy Consumption (Single-phase & Three Phase )
		AC Voltage(True RMS)
		Range : 0.0 to 600.0V
		Resolution : 0.1V
	AC woltzen	Accuracy: ± (1% rdg + 5 dgts) on 20 Hz to 100Hz
4	AC voltage	± (6% rdg + 5 dgts) on 100 Hz to 400Hz
		Input impedance : 1MW
		Effective measurement: 2V to 600 V
		Crest Factor : <2.3 @ 0 to 50V; < 1.8 @ 50 to 600 V  Overload Protection : 650 V AC / DC rms
		Overload Protection : 650 V AC / DC THIS
		AC Current (True RMS)
		Range: 0.0 to 1000 A
		Resolution: 0.1 A
_		Accuracy: ±(1.5% rdg + 10 dgts) on 50 Hz to 60 Hz
5	AC current	±(6% rdg + 10 dgts) on 45 Hz to 400 Hz Effective measurement : 2A to 1000A
		Crest Factor : <2.5 @ 0 to 100A; < 1.5 @ 100 to 1000 A
		Overload Protection : 1000A AC max. For 1 minute.
		Accuracy at the center of CT.
		D Management Assured
		Power Measurement Accuracy Based on sine wave input, power factor=1, 50 Hz/60Hz
		Effective Range: 4A to 1000A; 80V to 600V.
		Active Power(KW)
		Range : 0.00 to 600.0 KW
		Resolution: 0.01 KW
		Accuracy: ± (2.0% rdg + 5 dgts)
		Apparent Power(KVA)
	Power	Range: 0.00 to 600.0 KVA Resolution: 0.01 KVA
		Accuracy: ±(2.0% rdg + 5dgts)
6		Reactive power and power factor
		values obtained by calculation
		from active power , current and
	1	voltage measurements.
		Reactive Power (KVAR)
		Range: 0.00 to 600.0 KVAR Resolution: 0.01 KVAR
		Accuracy: ± (2.0% rdg + 5 dgts)
	95.	Power Factor(PF)
		Range: -1.000 to 0.000; 0.000 to 1.000
		Resolution: 0.001
		[Power factor]-1.000 to 1.000 [Frequency] 45.0 Hz to 400 Hz, PEAK, phase detection, max/min/avg value
		display, auto hold, electric meter comparison, unbalanced 3-phase power estimate display
		Harmonic measurement(level, percentage & THD%)  Power measurement of Active power (KW), Reactive power (KVAR), Apparent Power (KVA). Power Factor
		(PF) for 3P 4W, 3P 3W,1P 2W, 1 P 3W
		Triple display KW, V, A
8	Other functions	Over range indication
		Harmonic Measurement
		(Effective in the voltage and current)
		Measurement items: Level of each order, percentage of
		each order and total harmonic distortion (THD-F and THD-R) Measurement Range : Fundamental frequency 45Hz to 65Hz
		The source of the state of the
10	Safety	According to EN 61010-1, Protection class II ,Over voltage Category : CAT III 600V
12	Operating Temperature	O°C to 40°C, < 70 % R.H.
.,	Continue	5 A 150 A 2   500 A 1500 V 4.5
14	Crest factor	6 A/ 60 A range 3 or less, 600 A/ 600 V range 1.6 or less
16	Dustproof and	Atleast Grip: IP54 , Jaw: IP50
10	waterproof	
	Continuous operating	Minimum 20 Hours
	time (Battery Backup)	Lienda Manual Allientes Clin Cat. Test Die Cat Better in third & Commission
	Accessories Required	User's Manual, Alligator Clip Set , Test Pin Set, Battery installed & Carrying Case